




7100 - Pipettes, Transfer, Volumetric, Class A Accuracy, With Batch Certificate

- Complies with ISO / DIN 648
- Calibrated at 20°C
- Calibrated for delivery (TD, Ex)
- Generally used for accurate measurement and decanting of liquid
- Each pipette coded with unique no. supplied with calibration certificate giving exact volume with traceability to national standards

Product Code	Capacity ml	Tolerance ± ml	Quantity per Case
710001D	1	0.007	10
710002D	2	0.010	10
710005D	5	0.015	10
710006D	10	0.020	10
710008D	20	0.030	10
710009D	25	0.030	10
710012D	50	0.050	10
710016D	100	0.080	10

* Blue graduations can be supplied on request.

* Individual calibration certificate can be provided on request



7101 - Pipettes, Transfer, Volumetric, Class A, Accuracy As Per USP

- Complies with ASTM E-969-02
- Calibrated at 20°C
- Calibrated for delivery (TD, Ex)
- Generally used for accurate measurement and decanting of liquid
- Each pipette coded with unique no. supplied with calibration certificate giving exact volume with traceability to national standards

Product Code	Capacity ml	Tolerance ± ml	Quantity per Case
7101001	1	0.006	10
7101002	2	0.006	10
7101003	3	0.01	10
7101004	4	0.01	10
7101005	5	0.01	10
7101706	6	0.01	10
7101707	7	0.01	10
7101708	8	0.02	10
7101709	9	0.02	10
7101006	10	0.02	10
7101007	15	0.03	10
7101008	20	0.03	10
7101009	25	0.03	10
7101010	30	0.03	10
7101011	40	0.05	10
7101012	50	0.05	10
7101016	100	0.08	10

* Blue graduations can be supplied on request.

