

Cat. No	Description	Articles
---------	-------------	----------

“New-P.NIX® UP 900” Ultra Pure(UP) Water Purification System, Up-to 1.5L/min
 with Total Organic Carbon(TOC), 2-Steps of Filter Exchange Indicator, (UP) Up to 18.3MΩ·cm, **초순수 제조장치**



- **Made in Korea** ■ **Certified CE, ISO 9001, ISO 14001**
- **RS 232C Port and Volumetric Dispensing Function by Timer**
- **Auto Regular Recycling to Maintain the Optimum water Quality Level**
- **Programming by 2 Steps for the Filter Exchange**
- **Filter Replace Message Indicating by LCD Display & Blinking LED Light**
- * Microprocess One Touch Technology Controller
- * 16 × 4 Characters Backlight LCD Digital Display
- * Obtain GD, QD Mark, Awarded the Korean Excellent Products
- * Included 0.2 μ m Final Capsule Filter, UP Membrane Pack(Scholar/Bio)

Specifications & Ordering information

Spec.	Model	New-P.NIX UP 900			
	Scholar-type	UV-type	Bio-type	Integrate-type	
Certified Mark	CE, ISO 9001, ISO 14001				
Product Flow Rate(UP)	Max. 1.5L/min				
Product Water Quality	Up to 18.3MΩ·cm				
Water Quality Display	• UP Product : Up to 18.3MΩ·cm • Feed water : 0 to 250 μ S/cm • Temperature : 0 to 99.9℃				
Display	16 × 4 Characters Backlight LCD Digital Display				
Filter Composition	• 0.2 μ m Final Capsule Filter • UP Pack (DH.WWP.UPS)	• 0.2 μ m Final Capsule Filter • UP Pack (DH.WWP.UPU) • 254/185UV Lamp	• 0.2 μ m Final Capsule Filter • UP Pack (DH.WWP.UPS) • MW 5000 UF Filter(L)	• 0.2 μ m Final Capsule Filter • UP Pack(DH.WWP.UPU) • 254/185UV Lamp • MW 5000 UF Filter(L)	
Controller	Microprocess One Touch Technology				
Dimensions (w × d × h)	300 × 500 × h 470mm				
Power	210~230V, 50/60 Hz				
Cat. No.	DH.WWP.UPS	DH.WWP.UPU	DH.WWP.UP.B	DH.WWP.UPI	
Price, Each. ₩					



(1) White, HDPE Aspiration Bottle (2) Amber, HDPE Aspiration Bottle



(3) 0.2 μ m Final Capsule Filter



(4) P.NIX UP Pack, Scholar / Bio



(5) P.NIX UP Pack, Scholar-UV / Integrate

• TOC monitoring system is also available on request. (TOC monitoring display range : 0~200ppb)

Optional Accessories

		Ea.	Price, ₩
SL.Bot3102	(1) Bottle, HDPE, Aspiration, White, 5L, Neck id- ϕ 70, ϕ 200 × h360mm		
SL.Bot3103	, 10L, - ϕ 70, ϕ 250 × h460mm		
SL.Bot3104	, 20L, - ϕ 95, ϕ 300 × h555mm		
SL.Bot3105	, 50L, - ϕ 95, ϕ 380 × h700mm		
SL.Bot3106	(2) Bottle, HDPE, Aspiration, Amber, 5L, Neck id- ϕ 70, ϕ 200 × h360mm		
SL.Bot3107	, 10L, - ϕ 70, ϕ 250 × h460mm		
SL.Bot3108	, 20L, - ϕ 95, ϕ 300 × h555mm		
SL.Bot3109	, 50L, - ϕ 95, ϕ 380 × h700mm		

Spare Filters & UV Lamp

Cat. No.	Description	Filter Replace Period	Ea. Price, ₩
DH.WWP.0.2CF	(3) 0.2 μ m Final Capsule Filter	Approx. 6~8 Month	
DH.WWP.UPS	(4) P.NIX UP Pack, Scholar / Bio	Approx. 1 Year	
DH.WWP.UPU	(5) P.NIX UP Pack, Scholar-UV / Integrate		
DH.WWP.MWL	MW 5000 UF Filter(L)	Approx. 6~8 Month	
DH.WWP.UV	254/185 UV Lamp	Approx. 3 Year	

• The above mentioned Filter replacement period can be changed according to various factors such as feed water condition, producing volume and so on.